



# The Islamia University of Bahawalpur

## TENDER NOTICE

Sealed Tenders on item wise rate basis are invited from the Manufacturing Firms or their Authorized Distributors, Dealers and General order Suppliers having established credentials in terms of Technical, Financial and Managerial capabilities for the supply of **Equipment, Department of University College of Nursing**, The Islamia University of Bahawalpur on FOR basis.

Tender No.	Category (Quantity and specifications are available in the tender document)	Estimated Cost	5% Bid Security
A	<b>Purchase of Equipment Department of University College of Nursing</b>	<b>Rs.4,723,750/-</b>	<b>Rs.236,188/-</b>

- Interested bidders may get the Tender Documents containing detailed specifications, terms and conditions from the Purchase Section, Account Department Abbasia Campus (Old), the Islamia University of Bahawalpur during office hours on submission of written application on letter head and a copy of CNIC or can be down loaded from the IUB website [www.iub.edu.pk](http://www.iub.edu.pk) OR PPRA website [www.ppra.punjab.gov.pk](http://www.ppra.punjab.gov.pk) on payment of (non-refundable fee) of RS.1000/ (Separately for each category) through Bank challan HBL in A/C 14730000010403 or pay order in the name of the Treasurer, The Islamia University of Bahawalpur . Original receipt should be attached with Technical bid.
- 5% Bank Guarantee** of the estimated cost as mentioned above separately against each tender (Refundable) in the shape of CDR / Bank Guarantee. **The Original CDR shall be attached with the technical bid & copy of CDR should be attached with financial bid.**
- The rates should be quoted on FOR basis.
- Single Stage – Sigel Envelope** bidding procedure will be adopted
- Procurement shall be governed under the Punjab Procurement Rules, 2014.
- The rates should be quoted inclusive of all taxes. Copy of paid bank challan of professional tax and copy of the bid security shall be attached with the technical proposal.
- The Islamia University of Bahawalpur however, reserves the rights to reject all bids at any time prior to acceptance of a bid as per clause 35 of the Punjab Procurement Rules and grounds of rejection will be conveyed to the bidders upon their request.
- Tender should reach in the Office of the Secretary SMPC Accounts Division Old Campus, The Islamia University of Bahawalpur up to **21.02.2023 by 10:30AM**. Tenders will be opened on the same day at **11:00AM** in the presence of bidders or their representatives. In the office of Chairman Purchase Committee, Pharmacy department Khawaja Fareed Campus.
- Any bid submitted after the given time schedule shall not be accepted.
- All taxes will be applicable as per Govt. Rules & Regulations.
- The rates should be valid upto 120-days from the date of opening of the tender.
- In case of closed/forced holidays, tender opening time/date will be considered as the next working day.



Chairman SMPC  
The Islamia University of Bahawalpur  
Baghdad-ul-Jadeed Campus, Bahawalpur.  
Phone: 062-9255565

IPB-84  
Daily Pakistan Multan  
Daily Dawn Lahore



# The Islamia University of Bahawalpur

## PROCUREMENT OF EQUIPMENT FOR DEPARTMENT OF UNIVERSITY COLLEGE OF NURSING TENDER-A FORM (FINANCIAL BID)

1. Name of Firm: \_\_\_\_\_
2. Mailing Address: \_\_\_\_\_
3. Phone No: \_\_\_\_\_ 4. Fax No: \_\_\_\_\_
5. Tender Fee Challah No. & amount \_\_\_\_\_ 6.CDR No.& Amount \_\_\_\_\_
7. General Sale Tax No: \_\_\_\_\_ 8. Income Tax No. \_\_\_\_\_
9. Professional Tax No: \_\_\_\_\_

Sr	Name of Items	Quantity	Rate per item inclusive
01	<b>Advanced airway intubation training simulator</b> <ul style="list-style-type: none"> <li>• Advanced Airway Intubation Simulator</li> <li>• Storage case</li> <li>• Endotracheal intubation tube: 6.5 mm cuffed with stylet oral tube</li> <li>• 9-volt battery</li> <li>• 1-year limited warranty</li> <li>• Curriculum</li> </ul>	01	
02	<b>C.P.R manikin Child:</b> Clear anatomic characteristic, realistic touch feeling and life like skin color, vivid appearance; <ul style="list-style-type: none"> <li>• Simulate vital signs: squeeze the rubber ball to simulate brachinal artery pulse;</li> <li>• Can perform artificial respiratory, extracardial compression and airway open;</li> <li>• Operation mode: CPR training, can perform compression and inflation;</li> </ul>	01	
03	Respirators (Ambu Bag Silicone) Reduced fatigue during ventilation Easy to clean and sterilize Ergonomic, lightweight bag design	01	
04	Rebreather,	01	
05	<b>Non Rebreather</b> , The mask covers both your nose and mouth. One-way valves prevent exhaled air from reentering the oxygen reservoir.	01	
06	<b>Nebulizer Machine:</b> Compressed air neutralization• For treatment of upper and lower airways, colds, asthma and respiratory diseases <ul style="list-style-type: none"> <li>• Premium carrying bag • Optimal respirable dose delivery • Quick effective treatment</li> <li>• Support for wide voltage input</li> <li>• Powerful Device • Remove Chest Congestion</li> </ul>	01	
07	<b>Oxygen Cylinder</b> with Key (Capacity: 10liter Weight: 13kg Pressure: 1900 psiUsageTime:14hours aprox.(at1LPM) Condition: Imported, Brand New, Tested With Certificate)	01	
08	Oxygen Flow Meter	01	
09	<b>Height weight Scale</b> (Weighing 200kg,resolution 500g Platform size 28x 38.5cm 70 to 190cm height.)	01	
10	<b>Cardiac Monitor:</b> 12.1-inch high-resolution color TET Touch Screen Display. Standard configuration: SpO2, PR, NIBP, ECG, RESP, TEMP Thinner and lighter in design, convenient and portable for users. Support Wired and Wire less Network Real-time S-T segment analysis, pacemaker detection. Various interfaces: standard screen, trend screen, oxy CRG screen, NIBP list screen, big front screen.	01	

	<p>3-level audio/visual alarms.</p> <p>Efficient resistance to interference of defibrillator and HF knife.</p> <p>Up to 400 groups NIBP list, 6000seconds, ECG waveform recall, 60 alarm event records recall, 7-day trend chart in storage.</p> <p>Built-in rechargeable lithium battery, 2-hour working capability.</p> <p>Wireless central monitoring system applicable to ICU/CCU/OR</p>		
11	<p><b>Automated External Defibrillator (AED):</b> By lifting up the device cover, the starts up immediately and guides the rescuer step-by-step during the entire resuscitation process. The available either as semi-automatic or fully automatic device. In its automatic version, the delivers the shock without any action of the rescuer. In its semi-automatic version, the prompts the rescuer to deliver the shock by pressing the shock button.</p> <p><b>FEATURES</b></p> <p><b>Ease of use:</b> interface with 1-2-3 steps</p> <p><b>Pre-connected electrodes</b> for faster application to the patient's chest</p> <p><b>Automatic self-tests</b> for detection of electrode expiration and battery capacity</p> <p><b>Trilingual:</b> perfect for an international environment</p>	01	
12	<p><b>Nursing Skills Manikin ADULT</b>(Full-body, adult manikin withostomy,mouth care,Urethral Catheterization for male and femaleIrrigation of bladder for male andfemale Puncture of thorax, abdominalcavity, General care, bed bathes,dressing and undressing, offeringtraininginmanyclinicalproceduresfor standardpatientcare)</p>	01	
13	<p><b>Nursing Skills Manikin infants</b>(Full-body, adult manikin withostomy,mouth care,Urethral Catheterization for male and femaleIrrigation of bladder for male andfemale Puncture of thorax, abdominalcavity, General care, bed bathes,dressing and undressing, offeringtraininginmanyclinicalproceduresfor standardpatientcare)</p>	01	
14	<p><b>Medication/Crash-Cart::</b> 1. Built with steel and aluminum to increase durability and lifespan.2. Impact resistant Aluminum Bumper with color code function. 3. Securing all drawers right at the side with Central Key Locking System. 4. A smooth workflow facilitated by Auto-return Ball Bearing Drawer Slides. 5. Installing accessory at any height with Flexible Accessory Mount. 6. Extended working area with Slide-out Work Surface. 7. Breakaway- Lock or Lock Bar for single or all drawers</p>	01	
15	Bag Valve Mask infant with tubing	01	
16	Laryngoscope with all size blade infant and adults	01	
17	Stethoscope dual sided	01	
18	<b>Back caretray</b> (Imported Quality)		
19	Bath basin (Imported Quality)	1	
20	Basin trolley (Imported Quality)	1	
21	Patients kit (Gown and pajama for patient)	01	
22	Bed pan with cover (Imported Quality)		
23	Sphygmomanometer mercury with Stand &cuff and without stand	01	
24	Blanket ((Best Quality)		
25	Bedsidescreen ((Best Quality)		
26	Bedsidetable ((Best Quality)	01	
27	Drawsheet ((Best Quality)		
28	Pillow ((Best Quality)	01	
29	<p><b>Suction machine</b></p> <p>Adopting oil-free piston pump</p> <p>Clean-without pollution of oil and smoke</p> <p>Convenient-no need to add oil to maintain daily</p> <p>Safe-no plus pressure during using</p> <p>No flow backward pressure when the machine stops, so the liquid won't flow backward</p> <p>Technical Information Of Suction Machine 7A 23b</p> <p>Power Voltage: AC220V±22V,50Hz±1Hz</p> <p>Max negative pressure: ≥0.09MPa</p> <p>Noise:≤65dB (A)</p> <p>Power:≤180VA</p> <p><b>Max Pressure/ Pumping rate: ≥20L/min</b></p> <p>Reservoir capacity: 2500mL/pc, 2pieces</p>	01	
30	<p><b>Patient electronic Bed along with air Mattress</b></p> <p><b>Key Features</b></p> <p>Safety and Functionality at its core</p> <p>Exceptional value for money</p> <p>Four sectioned profiling mattress support</p> <p>Newly optimised mattress platform sections to reduce shear and friction</p>	01	

	Increased comfort when sitting Minimised risk of sliding downwards during profiling Suitable for a wide range of client groups* Wide range of side rails and side rail covers available		
31	<b>Baby Incubator: Technical specification:</b> Power supply: AC220V, 50Hz Power input: 600VA Air Temperature control range: 25° C~37° C Sensor precision: ≤ 0.3° C Changing of temperature: ≤ 0.5° C Temperature rising time: ≤ 30min Noise level: ≤ 55dB(A) Inclination of bassinet: 0~5° C Alarming: Audible and visible alarm indication, and it can indicate the real temperature and cut off the power supply of heater (except for the low deviation alarm), meanwhile, the skin temperature indication window can indicate the alarm code. For the alarm 1-6, the relevant light is on, and for the alarm 7, all alarming lights are on; All alarm sound can be cancelled except for the power failure alarm and the system alarm. Over-temp alarm: When the temperature ≤ 38° C; Upper Deviation alarm: Air temperature control 3.0° C Lower Deviation alarm: Air temperature control -3.0° C Sensor alarm: When the sensor is disconnected, open circuit, short circuit or put on the wrong place; Fan alarm: When the fan is blocked, speed ≤ 1000rpm or stop working; Power failure alarm: When there is no power or the power cord is disconnected. System alarm: When there is E2ROM malfunction inside of temperature controller. Air mode control, set temperature and air temperature are displayed separately; Natural air flow humidity; Failure alarm indication; RS-232 connector.	01	
32	<b>Wheel chair:</b> Chromed steel frame Fixed armrest and footrest 8"PVC castor, 24" solid wheel	01	
33	<b>Patient Stretcure Local:</b> With side gaurd.cylinder stand.mesh & heavy wheels.	01	
34	<b>Patient Stretcure; Description:</b> <ul style="list-style-type: none"> <li>• Made of high-strength aluminum alloy thick-wall pipes.</li> <li>• The bed surface is made of foamed cushion and the adjustable back making the wounded feel comfortable.</li> <li>• One person is enough to send the patient to the ambulance stretcher.</li> <li>• The mechanical foldaway legs are controlled by the handles on both ends of the stretcher. The legs can be folded under the stretcher automatically when loading into the vehicle.</li> </ul>	01	

**Note: All the bidders are required to mention the details of their product being offered by them in the Technical proposal on firm's letter head duly signed & stamped by owner of firm or its authorized representative and also clearly mention the name & relevant major specifications of their offered product while quoting the financial bid as well. On firms letter head duly signed & stamped Mere reproduction of our tender documents will not serve the purpose and the bid may be liable to be rejected by the procuring agency in such cases.**

## Terms & Conditions

1. All material supplied must be brand new strictly conforming to the given specifications .Old, reconditioned or refurbished equipment shall not be acceptable.
2. Only certified dealers/distributors of the following manufacturers are eligible: **A standard company from etc.**
3. Prices should be FOR. If the items supplied are not according to the required specification/make it will have to be replaced by the firms on their own cost.
4. Offer must be supported with comprehensive technical literature and specifications in original (English) for the materials/goods offered.
5. Validity of rate should be for 120 days from the date of opening of tenders.
6. The vendors will be responsible for any damages during Transit/Delivery. They will also be responsible for any accident and their consequent damages.
7. Each page of tender documents should be signed and stamped by the bidder and attached with technical bid.
8. Key Person Contact No.0345-8672094

**Signature:** \_\_\_\_\_  
**Name:** \_\_\_\_\_  
**Designation:** \_\_\_\_\_  
**Date:** \_\_\_\_\_

**Attachments:**

Earnest Money draft (Yes/No)

Affidavit for non-black listing (Yes/No)

Bid Validity (Yes/No)

Signed terms & conditions (Yes/No)

**Note:** Bid shall be signed by the bidder/authorized person for bidder.

# Technical Specification offered by the firm required letter head of the firm duly signed & stamped by Vender of Tender A

01	<p><b>Advanced airway intubation trainingsimulator</b></p> <ul style="list-style-type: none"> <li>• Advanced Airway Intubation Simulator</li> <li>• Storage case</li> <li>• Endotracheal intubation tube: 6.5 mm cuffed with stylet oral tube</li> <li>• 9-volt battery</li> <li>• 1-year limited warranty</li> <li>• Curriculum</li> </ul>
02	<p><b>C.P.R manikin Child:</b> Clear anatomic characteristic, realistic touch feeling and life like skin color, vivid appearance;</p> <ul style="list-style-type: none"> <li>• Simulate vital signs: squeeze the rubber ball to simulate brachinal artery pulse;</li> <li>• Can perform artificial respiratory, extracardial compression and airway open;</li> <li>• Operation mode: CPR training, can perform compression and inflation;</li> </ul>
03	Respirators (Ambu Bag Silicone) Reduced fatigue during ventilation Easy to clean and sterilize Ergonomic, lightweight bag design
04	Rebreather,
05	<b>Non Rebreather</b> , The mask covers both your nose and mouth. One-way valves prevent exhaled air from reentering the oxygen reservoir.
06	<b>Nebulizer Machine:</b> Compressed air neutralization• For treatment of upper and lower airways, colds, asthma and respiratory diseases • Premium carrying bag • Optimal respirable dose delivery • Quick effective treatment • Support for wide voltage input • Powerful Device • Remove Chest Congestion
07	<b>Oxygen Cylinder</b> with Key (Capacity: 10liter Weight: 13kg Pressure: 1900 psiUsageTime:14hours aprox.(at1LPM) Condition: Imported, Brand New, Tested With Certificate)
08	Oxygen Flow Meter
09	<b>Height weight Scale</b> (Weighing 200kg,resolution 500g Platform size 28x 38.5cm 70 to 190cm height.)
10	<p><b>Cardiac Monitor:</b> 12.1-inch high-resolution color TET Touch Screen Display. Standard configuration: SpO2, PR, NIBP, ECG, RESP, TEMP Thinner and lighter in design, convenient and portable for users. Support Wired and Wire less Network Real-time S-T segment analysis, pacemaker detection. Various interfaces: standard screen, trend screen, oxy CRG screen, NIBP list screen, big front screen. 3-level audio/visual alarms. Efficient resistance to interference of defibrillator and HF knife. Up to 400 groups NIBP list, 6000seconds, ECG waveform recall, 60 alarm event records recall, 7-day trend chart in storage. Built-in rechargeable lithium battery, 2-hour working capability. Wireless central monitoring system applicable to ICU/CCU/OR</p>
11	<p><b>Automated External Defibrillator (AED):</b> By lifting up the device cover, the starts up immediately and guides the rescuer step-by-step during the entire resuscitation process. The available either as semi-automatic or fully automatic device.In its automatic version, the delivers the shock without any action of the rescuer. In its semi-automatic version, the prompts the rescuer to deliver the shock by pressing the shock button. <b>FEATURES</b> <b>Ease of use:</b> interface with 1-2-3 steps <b>Pre-connected electrodes</b> for faster application to the patient's chest <b>Automatic self-tests</b> for detection of electrode expiration and battery capacity <b>Trilingual:</b> perfect for an international environment</p>
12	<b>Nursing Skills Manikin ADULT</b> (Full-body, adult manikin withostomy,mouth care,Urethral Catheterization for male and femaleIrrigation of bladder for male andfemale Puncture of thorax, abdominalcavity, General care, bed bathes,dressing and undressing, offeringtraininginmanyclinicalproceduresfor standardpatientcare)
13	<b>Nursing Skills Manikin infants</b> (Full-body, adult manikin withostomy,mouth care,Urethral Catheterization for male and femaleIrrigation of bladder for male andfemale Puncture of thorax, abdominalcavity,

	General care, bed bathes, dressing and undressing, offering training in many clinical procedures for standard patient care)
14	<b>Medication/Crash-Cart:</b> 1. Built with steel and aluminum to increase durability and lifespan. 2. Impact resistant Aluminum Bumper with color code function. 3. Securing all drawers right at the side with Central Key Locking System. 4. A smooth workflow facilitated by Auto-return Ball Bearing Drawer Slides. 5. Installing accessory at any height with Flexible Accessory Mount. 6. Extended working area with Slide-out Work Surface. 7. Breakaway- Lock or Lock Bar for single or all drawers
15	Bag Valve Mask infant with tubing
16	Laryngoscope with all size blade infant and adults
17	Stethoscope dual sided
18	<b>Back care tray</b> (Imported Quality)
19	Bath basin (Imported Quality)
20	Basin trolley (Imported Quality)
21	Patients kit (Gown and pajama for patient)
22	Bed pan with cover (Imported Quality)
23	Sphygmomanometer mercury with Stand & cuff and without stand
24	Blanket ((Best Quality)
25	Bedsidescreen ((Best Quality)
26	Bedsidetable ((Best Quality)
27	Drawsheet ((Best Quality)
28	Pillow ((Best Quality)
29	<b>Suction machine</b> Adopting oil-free piston pump Clean-without pollution of oil and smoke Convenient-no need to add oil to maintain daily Safe-no plus pressure during using No flow backward pressure when the machine stops, so the liquid won't flow backward Technical Information Of Suction Machine 7A 23b Power Voltage: AC220V±22V, 50Hz±1Hz Max negative pressure: ≥0.09MPa Noise: ≤65dB (A) Power: ≤180VA <b>Max Pressure/ Pumping rate: ≥20L/min</b> Reservoir capacity: 2500mL/pc, 2pieces
30	<b>Patient electronic Bed along with air Mattress</b> <b>Key Features</b> Safety and Functionality at its core Exceptional value for money Four sectioned profiling mattress support Newly optimised mattress platform sections to reduce shear and friction Increased comfort when sitting Minimised risk of sliding downwards during profiling Suitable for a wide range of client groups* Wide range of side rails and side rail covers available
31	<b>Baby Incubator: Technical specification:</b> Power supply: AC220V, 50Hz Power input: 600VA Air Temperature control range: 25° C~37° C Sensor precision: ≤ 0.3° C Changing of temperature: ≤ 0.5° C Temperature rising time: ≤ 30min Noise level: ≤ 55dB(A) Inclination of bassinet: 0~5° C Alarming: Audible and visible alarm indication, and it can indicate the real temperature and cut off the power supply of heater (except for the low deviation alarm), meanwhile, the skin temperature indication window can indicate the alarm code. For the alarm 1-6, the relevant light is on, and for the alarm 7, all alarming lights are on; All alarm sound can be cancelled except for the power failure alarm and the system alarm. Over-temp alarm: When the temperature ≤ 38° C; Upper Deviation alarm: Air temperature control 3.0° C Lower Deviation alarm: Air temperature control -3.0° C Sensor alarm: When the sensor is disconnected, open circuit, short circuit or put on the wrong place; Fan alarm: When the fan is blocked, speed ≤ 1000rpm or stop working; Power failure alarm: When there is no power or the power cord is disconnected. System alarm: When there is E2ROM malfunction inside of temperature controller. Air mode control, set temperature and air temperature are displayed separately; Natural air flow humidity; Failure alarm indication; RS-232 connector.
32	<b>Wheel chair:</b> Chromed steel frame Fixed armrest and footrest 8"PVC castor, 24" solid wheel
33	<b>Patient Stretcher Local:</b> With side guard, cylinder stand, mesh & heavy wheels.

**Patient Stretcure; Description:**

- Made of high-strength aluminum alloy thick-wall pipes.
- The bed surface is made of foamed cushion and the adjustable back making the wounded feel comfortable.
- One person is enough to send the patient to the ambulance stretcher.
- The mechanical foldaway legs are controlled by the handles on both ends of the stretcher.

The legs can be folded under the stretcher automatically when loading into the vehicle.