



The Islamia University of Bahawalpur

TENDER NOTICE

Sealed Tenders on item wise rate basis are invited from the Manufacturing Firms or their Authorized Distributors, Dealers and General order Suppliers having established credentials in terms of Technical, Financial and Managerial capabilities for the supply of **Lab Equipment's & Chemicals & Permanent Items for Veterinary College**, The Islamia University of Bahawalpur on FOR basis.

Tender No.	Category (Quantity and specifications are available in the tender document)	Estimated Cost	5% Bid Security
A	Purchase of Lab Equipment's for the Department of Parasitology	Rs. 483,000/-	Rs.24,150 /-
B	Purchase of Chemicals & Glassware for the Department of Epidemiology & Public Health	Rs. 240,000/-	Rs.12,000/-
C	Purchase of Equipment for the Department of Pathology	Rs. 854,000/-	Rs.42,700/-
D	Purchase of Laboratory Equipment's for the Department of Animal Breeding and Genetics	Rs. 400,000/-	Rs.20,000/-

- Interested bidders may get the Tender Documents containing detailed specifications, terms and conditions from the Purchase Section, Account Department Abbasia Campus (Old), the Islamia University of Bahawalpur during office hours on submission of written application on letter head and a copy of CNIC or can be down loaded from the IUB website www.iub.edu.pk OR PPRA website www.ppra.punjab.gov.pk on payment of (non-refundable fee) of Rs. 1000/ (Separately for each category) through Bank challan HBL in A/C 14730000010403 or pay order in the name of the Treasurer, The Islamia University of Bahawalpur . Original receipt should be attached with Technical bid.
- 5% Bank Guarantee** of the estimated cost as mentioned above separately against each tender (Refundable) in the shape of CDR / Bank Guarantee shall be attached with the bid, otherwise bid will be rejected.
- The rates should be quoted on FOR basis.
- Single Stage – Two Envelopes** bidding procedure will be adopted as per PPRA Rules 38- 2(a). The Envelopes shall be marked as “**TECHNICAL PROPOSAL**” AND “**FINANCIAL PROPOSAL**” in legible letters. Financial Proposal of bids found technically non-responsive shall be returned un-opened.
- Procurement shall be governed under the Punjab Procurement Rules, 2014.
- The rates should be quoted inclusive of all taxes. Copy of paid bank challan of professional tax and copy of the bid security shall be attached with the technical proposal.
- The Islamia University of Bahawalpur however, reserves the rights to reject all bids at any time prior to acceptance of a bid as per clause 35 of the Punjab Procurement Rules and grounds of rejection will be conveyed to the bidders upon their request.
- Tender should reach in the Office of the Secretary SMPC Accounts Division Old Campus, The Islamia University of Bahawalpur upto **13.12.2021 by 10:30 AM**. Tenders will be opened on the same day at **11:00AM** in the presence of bidders or their representatives. In the office of Chairman Purchase Committee, Pharmacy department Khawaja Fareed Campus.
- Any bid submitted after the given time schedule shall not be accepted.
- All taxes will be applicable as per Govt. Rules & Regulations.
- The rates should be valid upto 120-days from the date of opening of the tender.
- In case of closed/forced holidays, tender opening time/date will be considered as the next working day.



Chairman SMPC
The Islamia University of Bahawalpur
Baghdad-ul-Jadeed Campus, Bahawalpur.
Phone: 062-9255565



The Islamia University of Bahawalpur

PROCUREMENT OF LAB EQUIPMENT FOR PARASITOLOGY TENDER-A FORM (FINANCIAL BID)

1. Name of Firm: _____
2. Mailing Address: _____
3. Phone No: _____ 4. Fax No: _____
5. Tender Fee Challah No. & amount _____ 6. CDR No. & Amount _____
7. General Sale Tax No: _____ 8. Income Tax No. _____
9. Professional Tax No: _____

Sr	Name of Items	Quantity	Rate per item inclusive
01	Thermo Cycler Capacity: 96-well 3 zones VeriFlex Display Type: 8 in. Color TFT LCD Interface: USB Width (Metric): 24 cm Depth (Metric): 46 cm Height (Metric): 21 cm Width (English): 9.5 in. Depth (English): 18.1 in. Height (English): 8.3 in. Weight (Metric): 8.3 kg Weight (English): 18.3 lb. Voltage: 100/240 V Wattage: 600 w Block Configurations: 96-Well, 0.2 mL VeriFlexs Max. Ramp Rate: 4°C/Sec (Block), 3°C/Sec (Sample) Thermal Uniformity: <0.5°C (30 Sec After Reaching 95°C) Thermal Accuracy: ±0.0.25°C (35°C to 99.9°C) Thermal Range: 0 to 100°C For Use With (Application): PCR Product Type: Thermal Cycler Format: 0.2 mL tubes, 96-well plate Includes: SimpliAmp™ Thermal Cycler High-throughput Compatibility: Not High-throughput Compatible (Manual)	01	

Note: All the bidders are required to mention the details of their product being offered by them in the Technical proposal on firm's letter head duly signed & stamped by owner of firm or its authorized representative and also clearly mention the name & relevant major specifications of their offered product while quoting the financial bid as well. On firms letter head duly signed & stamped Mere reproduction of our tender documents will not serve the purpose and the bid may be liable to be rejected by the procuring agency in such cases.

Terms & Conditions

1. All material supplied must be brand new strictly conforming to the given specifications .Old, reconditioned or refurbished equipment shall not be acceptable.
2. Only certified dealers/distributors of the following manufacturers are eligible: **A standard company from etc.**
3. Prices should be FOR. If the items supplied are not according to the required specification/make it will have to be replaced by the firms on their own cost.
4. Offer must be supported with comprehensive technical literature and specifications in original (English) for the materials/goods offered.
5. Validity of rate should be for 120 days from the date of opening of tenders.
6. The vendors will be responsible for any damages during Transit/Delivery. They will also be responsible for any accident and their consequent damages.
7. Each page of tender documents should be signed and stamped by the bidder and attached with technical bid.
8. Key Person Contact No.0333-4875876

Signature: _____
Name: _____
Designation: _____
Date: _____

Attachments:

Earnest Money draft (Yes/No)
Affidavit for non-black listing (Yes/No)
Bid Validity (Yes/No)
Signed terms & conditions (Yes/No)

Note: Bid shall be signed by the bidder/authorized person for bidder.



The Islamia University of Bahawalpur

PROCUREMENT OF CHEMICALS FOR EPIDEMIOLOGY & PUBLIC HEALTH TENDER-B FORM (FINANCIAL BID)

4. Name of Firm: _____

5. Mailing Address: _____

6. Phone No: _____ 4. Fax No: _____

5. Tender Fee Challah No. & amount _____ 6.CDR No.& Amount _____

7. General Sale Tax No: _____ 8. Income Tax No. _____

9. Professional Tax No: _____

Sr	Name of Items	Packing	Quantity	Rate per item inclusive
01	Rappaport-Vassiliadis enrichment Broth	500 gm	01	
02	Tetrathionate broth (TTB)	500 gm	01	
03	Xylose Lysine Tergitol 4 (XLT-4) agar	500 gm	01	
04	Luria broth (LB)	500 gm	01	
05	Cooked meat medium	500 gm	01	
06	Muller-Hinton agar (MH) plates	500 gm	01	
07	Tryptic Soy Broth (TSB)	500 gm	01	
08	Perfringens agar	500 gm	01	
09	Selenite F broth	500 gm	01	
10	Salmonella Shigella (SS) agar	500 gm	01	
11	Triple Sugar Iron agar	500 gm	01	
12	Simmons Citrate agar	500 gm	01	
13	Sabouraud's Dextrose Agar	500 gm	01	
14	API20E, 20A Kits	02 (20 strips each)	02	
15	Antibiotic discs ampicillin (30µg), amoxicillin/clavulanic acid (10µg), amikacin (30µg), cefixime (5µg), ciprofloxacin (10µg), cefepime (30µg), clindamycin (10µg), erythromycin (30µg), enrofloxacin (30µg), gentamycin (30µg), imipenem (10µg), kanamycin (30µg), linezolid (30µg), nalidixic acid (30µg), oxacillin (5µg), tetracycline (30µg), trimethoprim/sulfamethoxazole (25µg), vancomycin (30µg).	250 disc	20	

Note: All the bidders are required to mention the details of their product being offered by them in the Technical proposal on firm's letter head duly signed & stamped by owner of firm or its authorized representative and also clearly mention the name & relevant major specifications of their offered product while quoting the financial bid as well. On firms letter head duly signed & stamped Mere reproduction of our tender documents will not serve the purpose and the bid may be liable to be rejected by the procuring agency in such cases.

Terms & Conditions

- All material supplied must be brand new strictly conforming to the given specifications .Old, reconditioned or refurbished equipment shall not be acceptable.
- Only certified dealers/distributors of the following manufacturers are eligible: **A standard company from etc.**
- Prices should be FOR. If the items supplied are not according to the required specification/make it will have to be replaced by the firms on their own cost.
- Offer must be supported with comprehensive technical literature and specifications in original (English) for the materials/goods offered.
- Validity of rate should be for 120 days from the date of opening of tenders.
- The vendors will be responsible for any damages during Transit/Delivery. They will also be responsible for any accident and their consequent damages.
- Each page of tender documents should be signed and stamped by the bidder and attached with technical bid.
- Key Person Contact No.062-9255567

Signature: _____

Name: _____

Designation: _____

Date: _____

Attachments:

Earnest Money draft (Yes/No)

Affidavit for non-black listing (Yes/No)

Bid Validity (Yes/No)

Signed terms & conditions (Yes/No)

Note: Bid shall be signed by the bidder/authorized person for bidder.



The Islamia University of Bahawalpur

PROCUREMENT OF EQUIPMENT FOR PATHOLOGY TENDER-C FORM (FINANCIAL BID)

7. Name of Firm: _____

8. Mailing Address: _____

9. Phone No: _____ 4. Fax No: _____

5. Tender Fee Challah No. & amount _____ 6.CDR No.& Amount _____

7. General Sale Tax No: _____ 8. Income Tax No. _____

9. Professional Tax No: _____

Sr	Name of Items	Quantity	Rate per item inclusive
01	Hematology Analyzer for Veterinary with stabilizer: 3 Part Differential WBC, LYM #, MID#, GRAN#, LYM%, MID%, GRAN %, RBC,HGB,HCT,MCV, MCHC, RW- CV, RDW-SD, PLT, MPC, PDW, PCT, P-LCC, Throughput: 60 Samples/hr, Sampling: 10µl Whole Blood, Pre-Diluted 20µl, Maximum Parameters: 21, Calibration and Control program for each animal, Storage: 100,000 samples with histogram, Patient Data entry: Both via keyboard and Touch Screen, Port: Rs232, Network and USB, Printer: 57mm internal thermal Printer, Specimen Mixer: external, Blood roller Mixer, Consumable: Regents, Calibrator, Controls (01 Set), Suitable for, Dog, Cat, HORSE, Pig, Cow, Buffalo, Rabbit, Monkey, Rats, Mics, Sheep, Goat, Camel, Another 03 types that user can define. (total 16 programs), Power Supply: 240V 50/60 Hz, Power 180 W	01	

Note: All the bidders are required to mention the details of their product being offered by them in the Technical proposal on firm's letter head duly signed & stamped by owner of firm or its authorized representative and also clearly mention the name & relevant major specifications of their offered product while quoting the financial bid as well. On firms letter head duly signed & stamped Mere reproduction of our tender documents will not serve the purpose and the bid may be liable to be rejected by the procuring agency in such cases.

Terms & Conditions

1. All material supplied must be brand new strictly conforming to the given specifications .Old, reconditioned or refurbished equipment shall not be acceptable.
2. Only certified dealers/distributors of the following manufacturers are eligible: **A standard company from etc.**
3. Prices should be FOR. If the items supplied are not according to the required specification/make it will have to be replaced by the firms on their own cost.
4. Offer must be supported with comprehensive technical literature and specifications in original (English) for the materials/goods offered.
5. Validity of rate should be for 120 days from the date of opening of tenders.
6. The vendors will be responsible for any damages during Transit/Delivery. They will also be responsible for any accident and their consequent damages.
7. Each page of tender documents should be signed and stamped by the bidder and attached with technical bid.
8. Key Person Contact No. 0333-7429439

Signature: _____

Name: _____

Designation: _____

Date: _____

Attachments:

Earnest Money draft (Yes/No)

Affidavit for non-black listing (Yes/No)

Bid Validity (Yes/No)

Signed terms & conditions (Yes/No)

Note: Bid shall be signed by the bidder/authorized person for bidder.



The Islamia University of Bahawalpur

PROCUREMENT OF LABORATORY EQUIPMENT FOR ANIMAL BREEDING AND GENETICS TENDER-D FORM (FINANCIAL BID)

9. Name of Firm: _____

10. Mailing Address: _____

11. Phone No: _____ 4. Fax No: _____

5. Tender Fee Challah No. & amount _____ 6.CDR No.& Amount _____

7. General Sale Tax No: _____ 8. Income Tax No. _____

9. Professional Tax No: _____

Sr	Name of Items	Quantity	Rate per item inclusive
01	Liquid nitrogen Container (10 liters) Capacity 10 Litters, should be of advanced country, Made up of stainless steel, Having six canasters according to container size, Six goblets of high quality that can bear the -196°C temperature, Goblets should be compatible to canaster size, Daily Evaporation (0.10L/Day), Static Holding Time (93 Day), Dynamic Holding Time (93 Days)	01	
02	Autoclave (vertical) High pressure steam sterilizer autoclave volume about 22 liters	01	
03	Lab constant temperature incubator (DH), Ce incubator High-quality cold rolling steel electrostatic spraying exterior, Stainless steel or cold rolling sheet inner chamber; Foursquare semicircle transition, convenient for cleaning; Space between shelves is adjustable, Micro-computer intelligent control system with PID program, large screen digital display and light touching operation key, Function of timing, over-temperature protection and auto-power-break-off, Double door seal, which is convenient to on/off	01	
04	Distillation apparatus and water storage small tanks Technical Specifications: Output: Pure water Type II (analytical grade) According ASTM D1193 (American Society for Testing & Material) Input: Tap water Production: 1 l/min (60liters/hour) Purification process control: By microprocessor installation requirements:12/220V AC +water supply (≤ 3m. distance / ≥ 2 bar Pressure) Purifies Tap water or Type II Analytical Grade Water, for general laboratory use.	01	

Note: All the bidders are required to mention the details of their product being offered by them in the Technical proposal on firm's letter head duly signed & stamped by owner of firm or its authorized representative and also clearly mention the name & relevant major specifications of their offered product while quoting the financial bid as well. On firms letter head duly signed & stamped Mere reproduction of our tender documents will not serve the purpose and the bid may be liable to be rejected by the procuring agency in such cases.

Terms & Conditions

- All material supplied must be brand new strictly conforming to the given specifications .Old, reconditioned or refurbished equipment shall not be acceptable.
- Only certified dealers/distributors of the following manufacturers are eligible: **A standard company from etc.**
- Prices should be FOR. If the items supplied are not according to the required specification/make it will have to be replaced by the firms on their own cost.
- Offer must be supported with comprehensive technical literature and specifications in original (English) for the materials/goods offered.
- Validity of rate should be for 120 days from the date of opening of tenders.
- The vendors will be responsible for any damages during Transit/Delivery. They will also be responsible for any accident and their consequent damages.
- Each page of tender documents should be signed and stamped by the bidder and attached with technical bid.
- Key Person Contact No.0333-6589227

Signature: _____
Name: _____
Designation: _____
Date: _____

Attachments:

Earnest Money draft (Yes/No)
Affidavit for non-black listing (Yes/No)
Bid Validity (Yes/No)
Signed terms & conditions (Yes/No)

Note: Bid shall be signed by the bidder/authorized person for bidder.

Technical Specification offered by the firm required
letter head of the firm duly signed & stamped by
Vender Of Tender A

01	<p>Thermo Cyclcr Capacity: 96-well 3 zones VeriFlex Display Type: 8 in. Color TFT LCD Interface: USB Width (Metric): 24 cm Depth (Metric): 46 cm Height (Metric): 21 cm Width (English): 9.5 in. Depth (English): 18.1 in. Height (English): 8.3 in. Weight (Metric): 8.3 kg Weight (English): 18.3 lb. Voltage: 100/240 V Wattage: 600 w Block Configurations: 96-Well, 0.2 mL VeriFlexs Max. Ramp Rate: 4°C/Sec (Block), 3°C/Sec (Sample) Thermal Uniformity: <0.5°C (30 Sec After Reaching 95°C) Thermal Accuracy: ±0.0.25°C (35°C to 99.9°C) Thermal Range: 0 to 100°C For Use With (Application): PCR Product Type: Thermal Cyclcr Format: 0.2 mL tubes, 96-well plate Includes: SimpliAmp™ Thermal Cyclcr High-throughput Compatibility: Not High- throughput Compatible (Manual)</p>
----	---

Technical Specification offered by the firm required
letter head of the firm duly signed & stamped by
Vender Of Tender B

01	Rappaport-Vassiliadis enrichment Broth
02	Tetrathionate broth (TTB)
03	Xylose Lysine Tergitol 4 (XLT-4) agar
04	Luria broth (LB)
05	Cooked meat mediun
06	Muller-Hinton agar (MH) plates
07	Tryptic Soy Broth (TSB)
08	Perfringens agar
09	Selenite F broth
10	Salmonella Shigella (SS) agar
11	Triple Sugar Iron agar
12	Simmons Citrate agar
13	Sabouraud's Dextrose Agar
14	AP120E, 20A Kits
15	Antibiotic discs ampicillin (30µg), amoxicillin/clavulanic acid (10µg), amikacin (30µg), cefixime (5µg), ciprofloxacin (10µg), cefepime (30µg), clindamycin (10µg), erythromycin (30µg), enrofloxacin (30µg), gentamycin (30µg), imipenem (10µg), kanamycin (30µg), linezolid (30µg), nalidixic acid (30µg), oxacillin (5µg), tetracycline (30µg), trimethoprim/sulfamethoxazole (25µg), vancomycin (30µg).

Technical Specification offered by the firm required
letter head of the firm duly signed & stamped by
Vender Of Tender C

01	<p>Hematology Analyzer for Veterinary with stabilizer:</p> <p>3 Part Differential WBC, LYM #, MID#, GRAN#, LYM%, MID%, GRAN %, RBC,HGB,HCT,MCV, MCHC, RW- CV, RDW-SD, PLT, MPC, PDW, PCT, P-LCC, Throughput: 60 Samples/hr, Sampling: 10µl Whole Blood, Pre-Diluted 20µl, Maximum Parameters: 21, Calibration and Control program for each animal, Storage: 100,000 samples with histogram, Patient Data entry: Both via keyboard and Touch Screen, Port: Rs232, Network and USB, Printer: 57mm internal thermal Printer, Specimen Mixer: external, Blood roller Mixer, Consumable: Regents, Calibrator, Controls (01 Set), Suitable for, Dog, Cat, HORSE, Pig, Cow, Buffalo, Rabbit, Monkey, Rats, Mics, Sheep, Goat, Camel, Another 03 types that user can define. (total 16 programs), Power Supply: 240V 50/60 Hz, Power 180 W</p>
----	--

Technical Specification offered by the firm required
letter head of the firm duly signed & stamped by
Vender Of Tender D

01	<p>Liquid nitrogen Container (10 liters) Capacity 10 Litters, should be of advanced country, Made up of stainless steel, Having six canasters according to container size, Six goblets of high quality that can bear the -196°C temperature, Goblets should be compatible to canaster size, Daily Evaporation (0.10L/Day), Static Holding Time (93 Day), Dynamic Holding Time (93 Days)</p>
02	<p>Autoclave (vertical) High pressure steam sterilizer autoclave volume about 22 liters</p>
03	<p>Lab constant temperature incubator (DH), Ce incubator High-quality cold rolling steel electrostatic spraying exterior, Stainless steel or cold rolling sheet inner chamber; Four square semicircle transition, convenient for cleaning; Space between shelves is adjustable, Micro-computer intelligent control system with PID program, large screen digital display and light touching operation key, Function of timing, over-temperature protection and auto-power-break-off, Double door seal, which is convenient to on/off</p>
04	<p>Distillation apparatus and water storage small tanks Technical Specifications: Output: Pure water Type II (analytical grade) According ASTM D1193 (American Society for Testing & Material) Input: Tap water Production: 1 l/min (60liters/hour) Purification process control: By microprocessor installation requirements: 12/220V AC +water supply ($\leq 3m$. distance / ≥ 2 bar Pressure) Purifies Tap water or Type II Analytical Grade Water, for general laboratory use.</p>